

-----  
Midmark M11 - v1.0.5  
Total Cycles: 17

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2021 10:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.1 F 0.0  
2:00 184.6 F 3.9  
4:00 206.6 F 9.6  
6:00 221.7 F 6.3  
8:00 235.6 F 11.5  
10:00 238.3 F 10.5  
12:00 239.9 F 10.7  
14:00 254.6 F 18.1  
16:00 267.0 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 273.0 F 29.4  
2:00 271.8 F 28.5  
3:00 272.7 F 29.6  
4:00 272.7 F 29.2  
  
Min 271.5 F 28.5  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 222.2 F 0.0  
10:00 226.8 F 0.0  
15:00 209.2 F 0.0  
20:00 202.8 F 0.0  
25:00 207.8 F 0.0  
30:00 197.5 F 0.0  
  
Drying Complete

FILLING: 2:16  
HEATING: 17:25  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:56:47  
CYCLE COMPLETE

04/26/2021 11:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01