
Midmark M11 - v1.0.0
Total Cycles: 4724

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2024 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.9 F 0.0
2:00 129.1 F 1.8
4:00 144.5 F 4.7
6:00 168.6 F 8.0
8:00 191.8 F 11.9
10:00 206.5 F 4.7
12:00 222.2 F 9.2
14:00 231.2 F 11.7
16:00 233.7 F 9.6
18:00 237.0 F 9.5
20:00 244.6 F 12.9
22:00 256.0 F 18.8
24:00 266.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.2 F 28.5
3:00 272.6 F 29.3
4:00 271.6 F 28.7

Min 271.2 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 206.5 F 0.0
10:00 217.7 F 0.0
15:00 225.5 F 0.0
20:00 228.8 F 0.0
25:00 228.6 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:41
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

07/12/2024 02:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____