
Midmark M11 - v1.0.5
Total Cycles: 2649

Name: RITTER2
Sterilizer ID: V2313228
07/14/2023 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.7 F 0.0
2:00 128.1 F 0.0
4:00 161.8 F 0.3
6:00 177.6 F 3.3
8:00 194.8 F 6.6
10:00 200.5 F 10.9
12:00 213.1 F 4.6
14:00 224.6 F 7.6
16:00 234.9 F 11.8
18:00 233.7 F 8.5
20:00 244.2 F 12.7
22:00 243.6 F 12.4
24:00 255.4 F 18.5
26:00 265.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.5
2:00 271.6 F 28.7
3:00 272.4 F 29.7
4:00 272.2 F 29.2

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 212.4 F 0.0
10:00 209.3 F 0.0
15:00 197.6 F 0.0
20:00 191.7 F 0.0
25:00 187.2 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:46
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:06:03
CYCLE COMPLETE

07/14/2023 04:40 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____