
Midmark M11 - v1.0.5
Total Cycles: 2687

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 169.4 F 1.4
4:00 186.3 F 5.2
6:00 202.2 F 9.4
8:00 220.6 F 10.1
10:00 225.6 F 8.4
12:00 237.0 F 13.2
14:00 238.0 F 10.6
16:00 236.1 F 9.1
18:00 249.7 F 15.5
20:00 260.6 F 21.7
22:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.1 F 29.5
2:00 272.2 F 29.1
3:00 271.4 F 29.0
4:00 272.5 F 29.5

Min 271.3 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 214.4 F 0.0
10:00 219.2 F 0.0
15:00 214.1 F 0.0
20:00 210.3 F 0.0
25:00 209.8 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

07/24/2023 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____