
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000827
14:49 17/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:44
STERILIZING 3:49
136.0 C 227kPa 3:49
136.3 C 229kPa 4:19
136.4 C 231kPa 4:49
136.2 C 231kPa 5:19
136.4 C 233kPa 5:49
136.1 C 229kPa 6:19
136.3 C 231kPa 6:49
136.3 C 232kPa 7:19
136.3 C 232kPa 7:19
VENTING 7:19
AIR DRYING 8:15
CYCLE COMPLETE 23:15

Digital Signature #
9E1C869E0A2A915D
