
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002571
8:52 13/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:41
STERILIZING 3:46
136.4 C 224kPa 3:46
136.4 C 225kPa 4:16
136.4 C 223kPa 4:46
136.3 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 224kPa 6:16
136.4 C 225kPa 6:46
136.1 C 223kPa 7:16
136.1 C 223kPa 7:16
VENTING 7:16
AIR DRYING 8:05
CYCLE COMPLETE 31:07

Digital Signature #
2B258AFF341260D
