
Midmark M11 - v1.0.5
Total Cycles: 589

Name: RITTER2
Sterilizer ID: V2313228
11/30/2021 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.6 F 0.0
2:00 148.5 F 1.2
4:00 153.1 F 3.7
6:00 176.6 F 7.6
8:00 198.9 F 12.4
10:00 217.4 F 6.4
12:00 230.9 F 11.4
14:00 231.7 F 8.4
16:00 240.6 F 11.4
18:00 244.4 F 13.2
20:00 256.5 F 19.6
22:00 267.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.7
2:00 272.1 F 30.0
3:00 271.8 F 29.3
4:00 272.4 F 30.1

Min 271.4 F 29.2
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 221.9 F 0.0
10:00 201.5 F 0.0
15:00 200.1 F 0.0
20:00 197.3 F 0.0
25:00 207.2 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

11/30/2021 03:19 PM
ACCEPTED: X
REJECTED:
Operator: U02