

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1179

Name: RITTER 1  
Sterilizer ID: V2178198  
09/16/2021 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.8 F 0.0  
2:00 197.9 F 3.2  
4:00 216.2 F 8.2  
6:00 226.0 F 10.6  
8:00 235.2 F 9.2  
10:00 238.1 F 9.5  
12:00 244.6 F 12.6  
14:00 242.1 F 11.2  
16:00 255.6 F 18.1  
18:00 267.2 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.1 F 28.3  
2:00 272.4 F 28.8  
3:00 271.8 F 28.0  
4:00 272.5 F 28.8  
  
Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
5:00 223.8 F 0.0  
10:00 213.0 F 0.0  
15:00 216.5 F 0.0  
20:00 211.4 F 0.0  
25:00 209.7 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:22  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:16  
CYCLE COMPLETE

09/16/2021 10:36 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02