

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006489  
12:57 17/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:03  
STERILIZING 2:58  
136.4 C 220kPa 2:58  
136.1 C 223kPa 3:28  
136.1 C 218kPa 3:58  
136.4 C 224kPa 4:28  
136.4 C 225kPa 4:58  
135.9 C 220kPa 5:28  
136.3 C 222kPa 5:58  
136.3 C 225kPa 6:28  
136.3 C 225kPa 6:28  
VENTING 6:28  
AIR DRYING 7:17  
CYCLE COMPLETE 7:40  
-----

Digital Signature #  
B0CB44DA14563535  
-----