

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3676

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2024 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.0 F 0.0  
2:00 141.6 F 1.5  
4:00 149.9 F 4.0  
6:00 174.7 F 7.4  
8:00 195.0 F 11.3  
10:00 212.5 F 4.3  
12:00 220.6 F 8.8  
14:00 231.8 F 13.3  
16:00 234.6 F 10.0  
18:00 236.9 F 9.5  
20:00 246.0 F 14.1  
22:00 257.1 F 20.1  
24:00 267.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.5 F 30.3  
2:00 272.1 F 29.7  
3:00 271.2 F 29.5  
4:00 272.2 F 30.3

Min 271.2 F 29.3  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
5:00 211.8 F 0.0  
10:00 222.9 F 0.0  
15:00 210.3 F 0.0  
20:00 205.8 F 0.0  
25:00 203.9 F 0.0  
30:00 203.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:37  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:04:18  
CYCLE COMPLETE

06/07/2024 10:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_