
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004526
10:47 14/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:40
STERILIZING 3:47
136.4 C 223kPa 3:47
136.4 C 226kPa 4:17
136.4 C 225kPa 4:47
136.4 C 224kPa 5:17
136.4 C 224kPa 5:47
136.4 C 224kPa 6:17
136.3 C 224kPa 6:47
136.4 C 225kPa 7:17
136.4 C 225kPa 7:17
VENTING 7:17
AIR DRYING 9:04
CYCLE COMPLETE 14:25

Digital Signature #
1B316B04D418F75B
