
Midmark M11 - v1.0.0
Total Cycles: 2151

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.6 F 0.0
2:00 134.3 F 0.0
4:00 144.8 F 1.4
6:00 181.2 F 5.3
8:00 197.4 F 9.7
10:00 213.3 F 7.7
12:00 227.7 F 8.0
14:00 239.0 F 12.4
16:00 234.7 F 9.0
18:00 241.8 F 10.4
20:00 245.1 F 13.0
22:00 256.9 F 19.2
24:00 267.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.5
3:00 272.3 F 29.2
4:00 272.4 F 28.9

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 234.9 F 0.0
10:00 234.0 F 0.0
15:00 227.3 F 0.0
20:00 224.6 F 0.0
25:00 221.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

07/19/2022 02:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____