

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000549  
14:21 16/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:19  
PRESSURIZING 3:23  
STERILIZING 4:36  
136.0 C 230kPa 4:36  
136.3 C 231kPa 5:06  
136.3 C 231kPa 5:36  
136.4 C 233kPa 6:06  
136.4 C 233kPa 6:36  
136.2 C 230kPa 7:06  
136.3 C 231kPa 7:36  
136.2 C 229kPa 8:06  
136.2 C 229kPa 8:06  
VENTING 8:06  
AIR DRYING 9:05  
CYCLE COMPLETE 24:05  
-----

Digital Signature #  
8A6375492E0FBDA8  
-----