

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.4uS / 1.5ppm  
CYCLE NUMBER 003696  
12:20 26/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:11  
STERILIZING 3:06  
136.0 C 220kPa 3:06  
136.4 C 224kPa 3:36  
136.1 C 222kPa 4:06  
136.4 C 224kPa 4:36  
136.4 C 223kPa 5:06  
136.3 C 220kPa 5:36  
136.4 C 224kPa 6:06  
136.3 C 224kPa 6:36  
136.3 C 224kPa 6:36  
VENTING 6:36  
AIR DRYING 7:26  
CYCLE COMPLETE 7:28

-----  
Digital Signature #  
19E9A29186B5768A  
-----