

-----  
Midmark M11 - v1.0.5  
Total Cycles: 188

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2021 05:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.0 F 0.0  
2:00 143.6 F 1.3  
4:00 163.9 F 4.1  
6:00 193.0 F 8.3  
8:00 211.0 F 13.4  
10:00 220.8 F 8.0  
12:00 234.4 F 13.5  
14:00 237.4 F 11.1  
16:00 236.8 F 9.6  
18:00 249.9 F 16.1  
20:00 261.1 F 22.5  
22:00 271.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 272.4 F 29.8  
2:00 271.5 F 30.0  
3:00 272.0 F 30.1  
4:00 271.9 F 30.1

Min 271.0 F 29.3  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.6 F 0.0  
5:00 221.4 F 0.0  
10:00 230.4 F 0.0  
15:00 225.0 F 0.0  
20:00 221.1 F 0.0  
25:00 216.4 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:35  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

07/12/2021 06:07 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01