
Midmark M11 - v1.0.5
Total Cycles: 2352

Name: RITTER2
Sterilizer ID: V2313228
05/11/2023 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 78.0 F 0.0
4:00 79.4 F 0.0
6:00 98.8 F 0.0
8:00 153.5 F 0.4
10:00 192.0 F 0.4
12:00 204.7 F 3.3
14:00 211.6 F 6.5
16:00 223.1 F 10.0
18:00 235.7 F 10.4
20:00 236.5 F 10.3
22:00 237.4 F 8.9
24:00 245.7 F 13.5
26:00 244.3 F 12.5
28:00 256.3 F 18.9
30:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.2 F 29.2
2:00 271.9 F 28.6
3:00 271.9 F 29.1
4:00 272.4 F 29.0

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 209.0 F 0.0
10:00 192.4 F 0.0
15:00 197.1 F 0.0
20:00 204.3 F 0.0
25:00 222.2 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

05/11/2023 08:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____