

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1049

Name: RITTER2  
Sterilizer ID: V2313228  
04/28/2022 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
2:00 188.6 F 4.0  
4:00 206.7 F 9.6  
6:00 220.7 F 6.9  
8:00 232.7 F 11.0  
10:00 234.3 F 8.9  
12:00 238.9 F 9.9  
14:00 248.7 F 15.0  
16:00 261.3 F 22.2  
18:00 272.7 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.7  
2:00 271.5 F 29.0  
3:00 272.6 F 29.9  
4:00 272.1 F 29.3  
  
Min 271.4 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 223.1 F 0.0  
10:00 231.2 F 0.0  
15:00 227.1 F 0.0  
20:00 224.9 F 0.0  
25:00 223.7 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 18:25  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:44  
CYCLE COMPLETE

04/28/2022 03:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_