
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002528
8:38 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:33
STERILIZING 3:41
136.4 C 223kPa 3:41
136.3 C 223kPa 4:11
136.3 C 223kPa 4:41
136.4 C 221kPa 5:11
136.3 C 222kPa 5:41
136.1 C 222kPa 6:11
136.2 C 223kPa 6:41
136.3 C 223kPa 7:11
136.3 C 223kPa 7:11
VENTING 7:11
AIR DRYING 7:59
CYCLE COMPLETE 8:26

Digital Signature #
3C274501D0568627
