
Midmark M11 - v1.0.0
Total Cycles: 2887

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 07:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.0 F 0.0
2:00 196.1 F 4.5
4:00 211.1 F 10.0
6:00 222.6 F 5.9
8:00 234.2 F 11.6
10:00 236.3 F 9.4
12:00 237.2 F 8.8
14:00 250.7 F 15.7
16:00 263.6 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.5 F 28.4
3:00 272.5 F 29.4
4:00 272.1 F 28.8

Min 271.3 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 225.0 F 0.0
10:00 232.5 F 0.0
15:00 228.9 F 0.0
20:00 228.0 F 0.0
25:00 230.1 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

02/09/2023 08:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____