
Midmark M11 - v1.0.0
Total Cycles: 2477

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 183.2 F 4.3
4:00 206.1 F 8.4
6:00 213.8 F 11.0
8:00 226.3 F 7.0
10:00 238.7 F 12.0
12:00 235.3 F 8.9
14:00 241.5 F 9.8
16:00 246.5 F 13.5
18:00 258.6 F 20.0
20:00 269.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.0
2:00 271.6 F 28.3
3:00 272.2 F 29.1
4:00 272.2 F 28.7

Min 271.3 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 234.9 F 0.0
10:00 239.4 F 0.0
15:00 229.6 F 0.0
20:00 227.8 F 0.0
25:00 230.5 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

10/24/2022 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____