
Midmark M11 - v1.0.0
Total Cycles: 3320

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.9 F 0.0
2:00 180.8 F 3.6
4:00 194.1 F 8.3
6:00 213.6 F 12.4
8:00 225.3 F 8.7
10:00 235.4 F 10.7
12:00 240.8 F 11.7
14:00 239.4 F 10.2
16:00 253.3 F 17.0
18:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.5 F 29.1
4:00 272.0 F 28.6

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 216.1 F 0.0
10:00 221.9 F 0.0
15:00 223.5 F 0.0
20:00 223.2 F 0.0
25:00 223.8 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

06/01/2023 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____