
Midmark M11 - v1.0.0
Total Cycles: 1705

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2022 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.7 F 0.0
2:00 141.5 F 0.3
4:00 152.5 F 2.0
6:00 175.3 F 4.7
8:00 205.4 F 8.8
10:00 230.7 F 13.3
12:00 228.4 F 7.6
14:00 239.9 F 12.5
16:00 237.4 F 9.5
18:00 237.9 F 9.1
20:00 247.7 F 13.9
22:00 259.5 F 20.4
24:00 270.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.5 F 28.7
2:00 271.0 F 28.0
3:00 272.4 F 28.8
4:00 271.4 F 28.0

Min 271.0 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 224.8 F 0.0
10:00 232.8 F 0.0
15:00 218.1 F 0.0
20:00 227.4 F 0.0
25:00 223.1 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:52
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

03/10/2022 10:30 AM
ACCEPTED: X
REJECTED:
Operator: U02