

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5337

Name: RITTER 1  
Sterilizer ID: V2178198  
02/12/2025 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.9 F 0.0  
2:00 116.8 F 1.3  
4:00 142.0 F 4.1  
6:00 165.9 F 7.6  
8:00 189.2 F 11.6  
10:00 212.2 F 3.1  
12:00 223.0 F 7.2  
14:00 235.5 F 11.3  
16:00 229.9 F 6.8  
18:00 243.1 F 12.5  
20:00 239.5 F 10.1  
22:00 252.3 F 16.4  
24:00 263.4 F 22.9  
26:00 272.0 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.7 F 28.8  
2:00 271.3 F 28.2  
3:00 272.6 F 28.8  
4:00 271.4 F 28.1

Min 271.2 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 200.5 F 0.0  
10:00 196.5 F 0.0  
15:00 204.8 F 0.0  
20:00 212.8 F 0.0  
25:00 229.8 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 26:09  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:04:22  
CYCLE COMPLETE

02/12/2025 02:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_