
Midmark M11 - v1.0.0
Total Cycles: 782

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2021 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.7 F 0.0
2:00 121.1 F 0.1
4:00 125.6 F 0.9
6:00 181.9 F 3.4
8:00 184.5 F 7.8
10:00 203.3 F 12.4
12:00 216.9 F 6.3
14:00 231.9 F 11.2
16:00 234.1 F 7.7
18:00 242.5 F 11.8
20:00 244.9 F 12.4
22:00 257.7 F 19.2
24:00 269.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.6
1:00 272.5 F 28.8
2:00 272.0 F 28.6
3:00 272.3 F 28.9
4:00 272.4 F 28.7

Min 271.8 F 28.5
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 232.7 F 0.0
10:00 234.1 F 0.0
15:00 227.3 F 0.0
20:00 218.2 F 0.0
25:00 178.9 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 25:32
STERILIZING: 4:00
VENTING: 2:60
DRYING: 26:06
TOTAL CYCLE: 00:59:28

04/14/2021 10:02 AM
ACCEPTED: X
REJECTED:
Operator: U01