
Midmark M11 - v1.0.5
Total Cycles: 2069

Name: RITTER2
Sterilizer ID: V2313228
02/21/2023 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 178.7 F 0.0
4:00 182.2 F 2.6
6:00 204.4 F 7.3
8:00 219.3 F 12.4
10:00 230.4 F 8.9
12:00 236.9 F 10.0
14:00 244.0 F 12.7
16:00 245.2 F 13.4
18:00 258.5 F 20.7
20:00 270.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.7
2:00 271.3 F 29.0
3:00 272.5 F 29.9
4:00 271.7 F 29.3

Min 271.1 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 220.0 F 0.0
10:00 232.4 F 0.0
15:00 229.6 F 0.0
20:00 229.7 F 0.0
25:00 229.2 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

02/21/2023 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____