
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001639
12:01 30/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 3:46
136.4 C 220kPa 3:46
136.2 C 223kPa 4:16
136.4 C 223kPa 4:46
136.0 C 221kPa 5:16
136.3 C 225kPa 5:46
136.2 C 221kPa 6:16
136.4 C 225kPa 6:46
136.3 C 223kPa 7:16
136.3 C 223kPa 7:16
VENTING 7:17
AIR DRYING 8:07
CYCLE COMPLETE 8:10

Digital Signature #
2D504BE03B4EF76C
