
Midmark M11 - v1.0.0
Total Cycles: 2780

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2023 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 77.2 F 0.0
4:00 92.1 F 0.0
6:00 116.0 F 0.0
8:00 149.5 F 1.4
10:00 175.3 F 4.0
12:00 192.6 F 7.1
14:00 206.9 F 10.5
16:00 218.6 F 6.9
18:00 225.0 F 7.4
20:00 235.0 F 11.6
22:00 233.5 F 8.0
24:00 245.3 F 12.8
26:00 243.5 F 12.0
28:00 255.2 F 18.1
30:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.5 F 29.1
4:00 271.6 F 28.3

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 219.3 F 0.0
10:00 226.6 F 0.0
15:00 230.4 F 0.0
20:00 227.3 F 0.0
25:00 230.1 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:48
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

01/13/2023 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____