
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003061
14:36 01/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:11
136.4 C 221kPa 4:11
136.5 C 223kPa 4:41
136.1 C 217kPa 5:11
136.3 C 223kPa 5:41
136.4 C 223kPa 6:11
136.0 C 222kPa 6:41
136.3 C 221kPa 7:11
136.4 C 218kPa 7:41
136.4 C 218kPa 7:41
VENTING 7:41
AIR DRYING 8:38
CYCLE COMPLETE 8:45

Digital Signature #
E410E3915D64FBC4
