
Midmark M11 - v1.0.5
Total Cycles: 3454

Name: RITTER2
Sterilizer ID: V2313228
03/13/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 76.8 F 0.0
4:00 102.4 F 0.0
6:00 110.6 F 1.3
8:00 141.6 F 3.9
10:00 171.8 F 6.9
12:00 187.4 F 10.8
14:00 206.2 F 4.6
16:00 216.1 F 7.2
18:00 227.7 F 11.2
20:00 227.8 F 7.1
22:00 241.2 F 12.6
24:00 240.3 F 10.9
26:00 252.8 F 17.2
28:00 263.9 F 23.8
30:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.3 F 29.5
2:00 271.4 F 29.2
3:00 272.3 F 29.4
4:00 271.3 F 29.2

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 213.7 F 0.0
10:00 200.2 F 0.0
15:00 197.2 F 0.0
20:00 202.7 F 0.0
25:00 206.2 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:07
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

03/13/2024 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____