

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1611

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2022 06:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.2 F 0.0  
2:00 78.4 F 0.0  
4:00 85.7 F 0.0  
6:00 116.5 F 0.1  
8:00 156.4 F 0.1  
10:00 186.3 F 0.5  
12:00 212.9 F 4.0  
14:00 224.6 F 7.5  
16:00 233.9 F 11.2  
18:00 228.3 F 7.0  
20:00 239.5 F 11.7  
22:00 235.7 F 9.4  
24:00 238.3 F 9.8  
26:00 247.3 F 14.4  
28:00 259.2 F 20.8  
30:00 269.6 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.5  
2:00 271.6 F 29.1  
3:00 272.5 F 29.5  
4:00 271.5 F 29.0

Min 271.4 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
5:00 183.9 F 0.0  
10:00 171.1 F 0.0  
15:00 170.7 F 0.0  
20:00 167.3 F 0.0  
25:00 165.8 F 0.0  
30:00 153.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 31:00  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:19  
CYCLE COMPLETE

10/18/2022 08:02 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02