

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2958

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2023 12:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
2:00 196.4 F 4.4  
4:00 210.0 F 10.4  
6:00 220.9 F 5.6  
8:00 235.8 F 11.6  
10:00 236.8 F 9.7  
12:00 237.7 F 9.1  
14:00 250.9 F 15.9  
16:00 263.6 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.0 F 28.5  
2:00 272.5 F 28.9  
3:00 272.4 F 29.3  
4:00 271.6 F 28.4  
  
Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 212.5 F 0.0  
10:00 233.7 F 0.0  
15:00 227.8 F 0.0  
20:00 221.1 F 0.0  
25:00 217.1 F 0.0  
30:00 216.3 F 0.0  
  
Drying Complete

FILLING: 2:30  
HEATING: 17:52  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:37  
CYCLE COMPLETE

10/02/2023 01:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_