
Midmark M11 - v1.0.5
Total Cycles: 2630

Name: RITTER2
Sterilizer ID: V2313228
07/11/2023 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 152.7 F 0.4
4:00 159.6 F 3.3
6:00 177.8 F 7.4
8:00 199.8 F 12.1
10:00 215.9 F 6.1
12:00 229.7 F 10.9
14:00 230.2 F 7.7
16:00 242.8 F 13.5
18:00 242.1 F 12.0
20:00 254.8 F 18.5
22:00 265.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.7
2:00 271.4 F 29.0
3:00 272.5 F 29.9
4:00 271.8 F 29.2

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 203.9 F 0.0
10:00 203.9 F 0.0
15:00 199.5 F 0.0
20:00 199.4 F 0.0
25:00 199.7 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:55
CYCLE COMPLETE

07/11/2023 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____