
Midmark M11 - v1.0.0
Total Cycles: 3476

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 07:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 71.6 F 0.0
4:00 99.1 F 0.0
6:00 108.5 F 0.6
8:00 137.6 F 2.2
10:00 163.5 F 4.5
12:00 185.8 F 7.8
14:00 200.6 F 11.5
16:00 212.2 F 3.5
18:00 222.9 F 7.0
20:00 232.5 F 10.9
22:00 230.7 F 7.4
24:00 241.1 F 11.8
26:00 237.3 F 9.4
28:00 250.2 F 15.5
30:00 261.0 F 21.6
32:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.3 F 29.1
4:00 272.3 F 28.8

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 203.4 F 0.0
10:00 216.5 F 0.0
15:00 228.0 F 0.0
20:00 228.8 F 0.0
25:00 225.8 F 0.0
30:00 223.6 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:47
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:10:46
CYCLE COMPLETE

07/04/2023 08:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____