
Midmark M11 - v1.0.5
Total Cycles: 3244

Name: RITTER2
Sterilizer ID: V2313228
12/21/2023 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.9 F 0.0
2:00 130.4 F 0.4
4:00 141.3 F 2.6
6:00 167.5 F 6.1
8:00 191.1 F 10.4
10:00 210.7 F 5.5
12:00 220.9 F 8.4
14:00 236.5 F 13.4
16:00 236.5 F 10.2
18:00 237.2 F 8.6
20:00 249.2 F 15.1
22:00 261.1 F 21.9
24:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.3 F 29.6
2:00 271.8 F 28.9
3:00 271.5 F 29.1
4:00 272.4 F 29.5

Min 271.1 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 230.8 F 0.0
10:00 195.0 F 0.0
15:00 185.2 F 0.0
20:00 184.6 F 0.0
25:00 210.7 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:37
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

12/21/2023 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____