

-----  
Midmark M11 - v1.0.5  
Total Cycles: 641

Name: RITTER2  
Sterilizer ID: V2313228  
12/16/2021 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.9 F 0.0  
2:00 139.2 F 1.7  
4:00 160.1 F 5.2  
6:00 189.7 F 10.1  
8:00 221.0 F 5.2  
10:00 226.3 F 9.7  
12:00 232.1 F 8.5  
14:00 241.4 F 12.4  
16:00 240.1 F 10.9  
18:00 252.8 F 17.6  
20:00 264.8 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.6 F 29.5  
2:00 271.9 F 29.7  
3:00 272.3 F 29.4  
4:00 271.8 F 29.7

Min 271.2 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.2 F 0.0  
5:00 226.2 F 0.0  
10:00 213.6 F 0.0  
15:00 229.2 F 0.0  
20:00 212.8 F 0.0  
25:00 218.4 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 21:52  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:02  
CYCLE COMPLETE

12/16/2021 10:06 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02