
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001595
14:55 11/18/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 7:06
STERILIZING 8:52
136.4 C 222kPa 8:52
136.1 C 220kPa 9:22
136.5 C 222kPa 9:52
136.2 C 220kPa 10:22
136.4 C 223kPa 10:52
136.4 C 223kPa 11:22
136.3 C 221kPa 11:52
136.2 C 221kPa 12:22
VENTING 12:22
AIR DRYING 13:32
CYCLE COMPLETE 14:59

Digital Signature #
E17E8F189112F6D8
