
Midmark M11 - v1.0.5
Total Cycles: 470

Name: RITTER2
Sterilizer ID: V2313228
10/21/2021 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.4 F 0.0
2:00 151.3 F 2.7
4:00 178.4 F 6.6
6:00 196.0 F 11.5
8:00 215.7 F 5.6
10:00 228.8 F 10.5
12:00 232.0 F 7.9
14:00 243.1 F 12.6
16:00 244.1 F 13.0
18:00 256.5 F 19.6
20:00 267.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 273.0 F 30.1
2:00 272.2 F 29.4
3:00 271.8 F 29.5
4:00 273.1 F 30.1

Min 271.6 F 29.1
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 221.7 F 0.0
10:00 219.2 F 0.0
15:00 225.2 F 0.0
20:00 218.9 F 0.0
25:00 217.6 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

10/21/2021 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____