
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001270
16:01 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:54
STERILIZING 5:05
136.0 C 221kPa 5:05
136.3 C 223kPa 5:35
136.3 C 224kPa 6:05
136.2 C 224kPa 6:35
136.4 C 222kPa 7:05
136.2 C 223kPa 7:35
136.3 C 225kPa 8:05
136.4 C 225kPa 8:35
136.4 C 225kPa 8:35
VENTING 8:35
AIR DRYING 9:30
CYCLE COMPLETE 24:30

Digital Signature #
39F0796E3EDF2960
