

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4793

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2024 08:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 73.0 F 0.0  
4:00 81.0 F 0.0  
6:00 115.7 F 0.2  
8:00 149.2 F 0.8  
10:00 177.8 F 2.3  
12:00 199.4 F 5.4  
14:00 210.3 F 8.9  
16:00 222.5 F 12.6  
18:00 224.8 F 6.3  
20:00 235.3 F 10.6  
22:00 234.1 F 7.7  
24:00 244.3 F 12.7  
26:00 241.6 F 10.9  
28:00 254.0 F 17.2  
30:00 265.0 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.7  
2:00 271.5 F 27.9  
3:00 272.8 F 28.8  
4:00 271.8 F 28.1  
  
Min 271.5 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 197.2 F 0.0  
10:00 187.4 F 0.0  
15:00 192.6 F 0.0  
20:00 221.5 F 0.0  
25:00 224.1 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 31:55  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:56  
CYCLE COMPLETE

08/08/2024 09:26 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_