
Midmark M11 - v1.0.0
Total Cycles: 746

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 152.0 F 1.1
4:00 177.4 F 4.4
6:00 190.5 F 9.4
8:00 216.9 F 5.8
10:00 222.3 F 8.3
12:00 235.0 F 13.5
14:00 238.1 F 10.7
16:00 235.4 F 8.6
18:00 250.1 F 15.5
20:00 262.0 F 22.4
22:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.1 F 28.7
2:00 272.1 F 29.1
3:00 272.1 F 29.1
4:00 271.8 F 28.6

Min 271.2 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 233.3 F 0.0
10:00 221.9 F 0.0
15:00 223.7 F 0.0
20:00 224.2 F 0.0
25:00 228.0 F 0.0
30:00 232.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:13
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

03/25/2021 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____