

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4875

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2024 03:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.8 F 0.0  
2:00 158.7 F 0.3  
4:00 182.2 F 1.1  
6:00 205.0 F 4.6  
8:00 221.4 F 8.5  
10:00 227.7 F 12.8  
12:00 230.8 F 7.6  
14:00 242.7 F 12.6  
16:00 239.6 F 10.3  
18:00 236.6 F 8.4  
20:00 251.0 F 15.5  
22:00 262.4 F 22.0  
24:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.0  
1:00 272.1 F 28.6  
2:00 272.3 F 28.5  
3:00 271.5 F 28.0  
4:00 272.4 F 28.9

Min 271.4 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 200.1 F 0.0  
10:00 201.0 F 0.0  
15:00 201.0 F 0.0  
20:00 239.3 F 0.0  
25:00 239.3 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:03:06  
CYCLE COMPLETE

09/06/2024 04:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_