
Midmark M11 - v1.0.0
Total Cycles: 4790

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.1 F 0.0
2:00 110.0 F 0.1
4:00 139.6 F 0.8
6:00 168.7 F 3.0
8:00 182.8 F 5.9
10:00 225.7 F 9.5
12:00 219.5 F 13.3
14:00 223.9 F 6.4
16:00 234.5 F 10.5
18:00 234.0 F 8.1
20:00 242.3 F 11.6
22:00 238.1 F 9.4
24:00 250.6 F 15.3
26:00 261.1 F 21.2
28:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.5 F 29.0
2:00 272.2 F 28.5
3:00 271.1 F 28.0
4:00 272.4 F 29.1

Min 271.0 F 28.0
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 199.3 F 0.0
10:00 186.6 F 0.0
15:00 186.2 F 0.0
20:00 185.5 F 0.0
25:00 184.6 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:49
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:07:29
CYCLE COMPLETE

08/07/2024 03:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____