
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001083
14:23 18/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:52
STERILIZING 4:57
136.2 C 220kPa 4:57
136.4 C 225kPa 5:27
136.4 C 225kPa 5:57
136.3 C 224kPa 6:27
136.3 C 224kPa 6:57
136.3 C 222kPa 7:27
136.4 C 225kPa 7:57
136.4 C 223kPa 8:27
136.4 C 223kPa 8:27
VENTING 8:27
AIR DRYING 9:22
CYCLE COMPLETE 24:22

Digital Signature #
E1B3C075965DABC9
