

-----  
Midmark M11 - v1.0.0  
Total Cycles: 934

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2021 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	74.2 F	0.0
2:00	93.4 F	0.0
4:00	91.8 F	0.0
6:00	117.8 F	0.0
8:00	146.1 F	0.1
10:00	172.3 F	0.2
12:00	189.6 F	2.2
14:00	208.4 F	4.5
16:00	219.4 F	7.3
18:00	227.7 F	10.5
20:00	232.7 F	9.5
22:00	235.3 F	9.1
24:00	242.5 F	12.1
26:00	242.2 F	11.4
28:00	238.8 F	9.5
30:00	251.2 F	15.6
32:00	261.6 F	21.5
34:00	271.4 F	27.9

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.0
1:00	272.6 F	28.7
2:00	271.8 F	28.0
3:00	271.9 F	28.5
4:00	272.5 F	28.6

Min 271.3 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	210.4 F	0.0
5:00	210.5 F	0.0
10:00	197.5 F	0.0
15:00	195.6 F	0.0
20:00	193.0 F	0.0
25:00	194.7 F	0.0
30:00	203.4 F	0.0

Drying Complete

FILLING: 0:54  
HEATING: 34:44  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:12:58  
CYCLE COMPLETE

06/24/2021 09:33 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_