

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3797

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2023 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 69.7 F 0.0  
4:00 74.6 F 0.0  
6:00 100.4 F 0.0  
8:00 138.6 F 0.9  
10:00 169.2 F 2.6  
12:00 191.2 F 4.6  
14:00 208.3 F 7.8  
16:00 211.0 F 11.2  
18:00 219.8 F 5.4  
20:00 228.2 F 8.1  
22:00 236.2 F 12.2  
24:00 235.2 F 8.6  
26:00 244.3 F 12.1  
28:00 244.1 F 12.1  
30:00 255.6 F 18.0  
32:00 265.7 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.6 F 28.7  
2:00 271.5 F 27.9  
3:00 272.4 F 28.8  
4:00 272.1 F 28.3  
  
Min 271.3 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 237.2 F 0.0  
10:00 218.1 F 0.0  
15:00 198.5 F 0.0  
20:00 191.3 F 0.0  
25:00 193.6 F 0.0  
30:00 193.2 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 33:53  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:11:59  
CYCLE COMPLETE

09/20/2023 09:40 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_