
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002613
11:31 03/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:50
136.4 C 224kPa 3:50
136.2 C 221kPa 4:20
136.2 C 218kPa 4:50
136.3 C 223kPa 5:20
136.4 C 221kPa 5:50
136.4 C 224kPa 6:20
136.3 C 222kPa 6:50
136.4 C 221kPa 7:20
136.4 C 221kPa 7:20
VENTING 7:20
AIR DRYING 8:14
CYCLE COMPLETE 8:15

Digital Signature #
CDC5C78A14228194
