
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006701
12:05 06/12/2023
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 4:26
STERILIZING 5:21
136.0 C 220kPa 5:21
136.4 C 218kPa 5:51
136.2 C 222kPa 6:21
136.4 C 219kPa 6:51
136.4 C 222kPa 7:21
136.4 C 222kPa 7:51
136.2 C 221kPa 8:21
136.4 C 224kPa 8:51
136.4 C 224kPa 9:21
136.4 C 222kPa 9:51
136.2 C 218kPa 10:21
136.4 C 223kPa 10:51
136.4 C 224kPa 11:21
136.3 C 219kPa 11:51
136.4 C 216kPa 12:21
136.2 C 222kPa 12:51
136.2 C 218kPa 13:21
136.3 C 223kPa 13:51
136.1 C 220kPa 14:21
136.1 C 221kPa 14:51
136.4 C 223kPa 15:21
136.4 C 223kPa 15:21
VENTING 15:21
AIR DRYING 16:11
CYCLE COMPLETE 16:15

Digital Signature #

5D07AB7CB0BEB098
