

-----  
Midmark M11 - v1.0.0  
Total Cycles: 716

Name: MIDMARK  
Sterilizer ID: V2178198  
03/02/2021 07:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.2 F 0.0  
2:00 74.5 F 0.0  
4:00 79.7 F 0.0  
6:00 106.0 F 0.1  
8:00 153.0 F 0.3  
10:00 187.9 F 2.9  
12:00 205.3 F 6.3  
14:00 217.1 F 10.0  
16:00 230.7 F 9.3  
18:00 235.3 F 9.6  
20:00 235.6 F 8.9  
22:00 245.4 F 12.8  
24:00 242.6 F 11.3  
26:00 256.7 F 18.5  
28:00 268.6 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 27.9  
1:00 272.6 F 28.6  
2:00 271.3 F 27.9  
3:00 272.2 F 28.3  
4:00 271.8 F 28.4

Min 271.1 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 199.1 F 0.0  
10:00 200.5 F 0.0  
15:00 197.7 F 0.0  
20:00 193.8 F 0.0  
25:00 191.0 F 0.0  
30:00 189.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 28:58  
STERILIZING: 4:00  
VENTING: 2:50  
DRYING: 30:00  
TOTAL CYCLE: 01:06:43  
CYCLE COMPLETE

03/02/2021 08:50 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_