

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1429

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2022 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 74.9 F 0.0  
4:00 80.1 F 0.0  
6:00 102.4 F 0.0  
8:00 143.8 F 1.6  
10:00 174.0 F 3.5  
12:00 199.7 F 6.2  
14:00 227.9 F 9.9  
16:00 235.1 F 11.0  
18:00 232.9 F 8.9  
20:00 240.4 F 12.8  
22:00 238.5 F 11.1  
24:00 237.0 F 9.3  
26:00 250.1 F 15.7  
28:00 261.0 F 22.0  
30:00 271.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 273.0 F 29.8  
2:00 271.5 F 28.9  
3:00 272.2 F 29.7  
4:00 272.6 F 29.4  
  
Min 271.0 F 28.8  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
5:00 201.9 F 0.0  
10:00 182.0 F 0.0  
15:00 179.2 F 0.0  
20:00 166.6 F 0.0  
25:00 158.9 F 0.0  
30:00 152.4 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 30:46  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

08/30/2022 07:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_