
Midmark M11 - v1.0.5
Total Cycles: 3221

Name: RITTER2
Sterilizer ID: V2313228
12/15/2023 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.3 F 0.0
2:00 76.8 F 0.0
4:00 85.7 F 0.4
6:00 119.6 F 2.0
8:00 153.0 F 4.9
10:00 180.8 F 8.1
12:00 213.6 F 12.4
14:00 219.7 F 5.3
16:00 228.6 F 9.0
18:00 235.2 F 13.1
20:00 234.9 F 9.0
22:00 240.5 F 11.3
24:00 243.5 F 12.2
26:00 254.6 F 18.0
28:00 264.8 F 24.2
30:00 271.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.3 F 29.1
2:00 271.4 F 29.0
3:00 271.9 F 28.9
4:00 271.8 F 29.3

Min 271.0 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 243.8 F 0.0
10:00 230.9 F 0.0
15:00 221.5 F 0.0
20:00 233.4 F 0.0
25:00 229.4 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:01
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:06
CYCLE COMPLETE

12/15/2023 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____